

GOVERNMENT POLYTECHNIC, DHULE.



Vidyanagri Deopur, Dhule

PH: 02562-272103. FAX: 02562-272103.

Email ID :- stores.gpdhule@dtmaharashtra.gov.in

office.gpdhule@dtmaharashtra.gov.in

Website : gpdhule.org.in

No. : GPD/STORES/INST.PURC-AUTO-B2/2017-18 /35)

Date:- 3 FEB 2018

QUOTATION

To,

**Sub: Quotation for purchase of equipment.
(Due Date:- 09.02.2018)**

Sealed Quotations are from eligible and interested manufacturers/ dealers/distributors/ for the following items on the terms and conditions mentioned below.

TERMS AND CONDITIONS

- The bid/quotation should be submitted in two bid format.
- The first envelope will contain Technical specifications of the product, technical literature/ leaflet and other documents mentioned below:-
 - Covering Letter for tender on the company letter head mentioning official address, Contact No, e Mail address and website (if available) address
 - GST registration certificate/ Number
 - GST Clearance Certificate/ GST Challan for last quarter of the financial year.
 - Authorization/ Distributorship certificate from manufacturer. Proof of permission to manufacture the equipment/ item mentioned in the quotation from competent authorities (to be submitted if the bidder is not a manufacturer).
 - Technical literature / leaflet of the make and model no of equipment quoted.Additional document may also be asked by undersigned for confirming the details.
- The second envelope will contain the financial bid in which the all-inclusive rates F.O.R. Institute G.P.Dhule will be written and signed with the stamp of the establishment in the following format.

Sr. No.	Name of the item with specification	Quantity required	All inclusive cost per unit	All inclusive cost for quantity mentioned.
1.				

Date: -
Place: -

(Signature)
Rubber Stamp of organization.

- The two envelopes should be sealed with a mention of the type of envelope (technical/ Financial), Reference no., Date of opening the quotation on the front side of the envelope. These two envelopes should be sealed in a third envelope by giving heading **"Quotation for supply of AUTOMOBILE EQUIPMENTS"** and writing complete address of the undersigned.
- The quotations should reach the undersigned on or before date **09.02.2018**
- The material will be checked at this institute.
- Prices should be quoted F.O.R Institute (Govt.Polytechnic Dhule) and no extra charges will be paid for cartage, packing etc. for the material rejected and replaced
- Quotations will be opened at **11:00 AM** on date **10.02.2018**
- Rates should be valid for 6 months from the date of confirmation letter.
- Materials should be quoted for standard makes and minimum pkgs.
- Delivery to the consignee has to be effected within 02 weeks from the date of issue of purchase order. Also the undersigned reserves the rights to reject the same if the delivery of the stores is delayed beyond the stipulated period.
- Immediately after delivery of goods, the supplier must send his team of expert for Demonstration/ Installation/ Commissioning/Erection/Testing and Checking which shall be carried out free of cost at Govt.Polytechnic, Dhule and also the material required for testing should be carried for demonstration.
- Prior information for delivery of goods as well as for Demonstration/ Installation/ Commissioning/Erection/Testing and checking should be communicated .
- No advance payment is permissible.
- Every case will be taken to clear up the bills in a lease possible period, if acceptable material to the agreed specifications mentioned in the purchase order must be supplied immediately without any delay.
- Dispatched documents must be sent directly to this office and not through any bankers.
- Material will be strictly rejected if not found as per purchase order specifications and the same will be returned to your firm at your own cost.
- Warranty clause must be mentioned separately in quotation. The supplier shall ensure that all service calls will be attended by him/his authorized service provider promptly within the warranty period and even after the warranty period is over.
- The undersigned reserves the right to alter /amend any terms without prior reference to the supplier and ordering part quantities without assigning reasons and the supplier shall supply the ordered quantity at accepted rates.
- The undersigned reserves the right to accept or reject any offer or all offers without assigning any reason thereof.


(R.K.CHAUDHARI)

PRINCIPAL
GOVT.POLYTECHNIC,DHULE

COPY SUBMITTED TO :

1. THE DIRECTOR ,DIRECTORATE OF TECHNICAL EDUCATION – WEB PORTAL
2. THE JOINT DIRECTOR ,REGIONAL OFFICE NASHIK – WEB PORTAL
3. GOVT.POLYTECHNIC,DHULE INSTITUTE WEB PORTAL
4. NOTICE BOARD, GOVT.POLYTECHNIC,DHULE
5. OFFICE SECTION :- TO BE PUBLISHED AT LOCAL NEWSPAPER AT IMMEDIATE PRIORITY BASIS.

GOVERNMENT POLYTECHNIC, DHULE.



Vidyanagri Deopur ,Dhule

PH: 02562-272103. FAX: 02562-272103.


Email ID :- stores.gpdhule@dtmaharashtra.gov.in

office.gpdhule@dtmaharashtra.gov.in

Website : gpdhule.org.in

Priority no	Name of the Equipment to be procured in Detail specification in generic	Required Quantity
11	Vehicle Starter Motor Voltage- 12V, Power -0.9 KW to 1.4 KW	1
12	Working Model of Hydraulic Brake System two disc brakes in the front and two drum brakes at the rear and it is to be made out of original parts such as two master cylinder Assembly (one sectioned and one working), wheel cylinder, Brake drum, Brake Shoes, calipers etc., suitably sectioned and mounted on a sturdy iron frame. By operating the lever provided, the working procedure of the model can be demonstrated	1
13	Vehicle Alternator 12V Alternator 12v 90 Amp	1
14	Digital Clamp Meter Digital AC/DC Clamp Meter, Digits LCD Display, Jaw Opening Size 25 mm, One Pair of test leads , Instruction Manual , Carrying Case , Battery Installed , K Type Thermo Couple	1
15	Hydraulic Trainer kit. Basic component such as--- Valves(3/2 and 5/2,) Profile plate , Pressure guage assembly , Dual Pressure valve, shuttle valve, Flow Controll valve , Cylinder (Single and double acting), Filter Regulator with guage, Manifold Assembly, Pressure Regulator, etc.	1
16	Turbocharger . Turbocharger employed on either diesel 4 wheeler vehicle or diesel power generator set	2
17	Variomatic transmission of four wheeler Cut section model of Variomatic transmission of four wheeler	1
18	Single Stage Reciprocating air Compressor. Working pressure at outlet: 09 Kg/Cm2 Receiver Air tank capacity: 300 liters. powered by a 3 phase A.C., 3 HP electric motor Piston Displacement:360lpm FAD:256 lpm	1
19	Orifice meter Tes Rig Apparatus to determine of CD,CV, CC,OF Orifice with accessrios like collection tank, stand, scale and slide attachment.	1
20	Venturimeter Test Rig. Venturimeter (cast iron /Brass) 25mm and 50mm.	1
21	General tools. Open ended spanner set, ring spanner, socket spanner, speed handle socket driver, rachet rive, extension, adjustable wrench, mallets, allen key sets, pliers-combination, nose, internal & external bend type, monkey plier, snap plier, bend nose plier	3
22	Tool kit for Auto airconditioning system servicing	2

23	Centrifugal Governor Test Rig Drive Unit- D.C. motor 0.25 hp. 0-1500 RPM • Speed Control Unit • Belt and Pulley system to give spindle speed 1 00 to 500 rpm. • Governor mechanisms: - Watt - Porter - Hartnell - Proell • Set of weights • Springs: 2 nos. of different stiffness of same length.	1
24	Hydraulic Trolley Type GARAGE JACK Capacity:- 3 Ton Minimum Height:- 135 mm Maximum Height:- 495 mm Overall Size(L*W*H) :- 650 X 340 X 180 mm Net Weight:-33.5	1
25	Oil dispenser	1
26	Working Model of Exhaust & Emmision System with EGR & PCV Valve	1
29	Professional Multimeter DC Volts: 200mV - 1000mV AC Volts: 2V - 750V DC Amps: 200mA - 10A OHMS: 200Ω - 20MΩ Dwell: 4, 6, 8 cylinder Tests starting & charging, ignition and fuel delivery systems, Locate faulty wiring and components Extra-large, easy-to-read digital LCD On-screen test leads connection reminder Automatic reverse polarity indication, Automatic zero adjust, Overload protection on all ranges Includes protective boot, detachable color-coded test leads, alligator clips and detailed instructions, Auto reverse polarity, Auto zero adjust, Overload protection, Low battery indicator, 10M Input, 9-Volt power, Audible tone, Auto power off, Includes diode check, Includes tach function, Includes continuity function.	1
30	Digital Bomb Calorimeter • Temperature Indication:- 3 ½ Digit - Digital Display • Differential Temperature Range:- 0 - 100 deg C • Reading Accuracy of Temperature:-0.01deg C • Firing Voltage :- 12 V AC • LED Indication for filament continuity • LED Indication for Bomb Calorimeter Combustion • Differential Digital Temperature Indicator with Built-in Timer with Computer Interface & Software	1


 (R.K. CHAUDHARI)
 PRINCIPAL
 GOVT. POLYTECHNIC, DHULE